



Calibration Laboratory Cert: 5518.01

ISO/IEC 17025:2017 and ANSI/NCSL Z540.1-1994

## Accredited Calibration Certificate

Customer:

Order / RMA:

Certificate:

Product: ESD Simulator

Manufacturer: TESEQ

Model: NSG438A

Serial: 187

Notes: 150pF / 330Ω INA 4380 - S/N: 38A187

Report Issued: 05/29/2025

Date of Calibration: 05/29/2025

Next Calibration:

*The next calibration date is defined by the equipment user/owner.*

The above instrument was tested and found to be within the Manufacturer's specifications at the tested parameters. The results of the tests performed are held on file at The EMC Shop. The calibration was carried out in accordance with the general requirements of ISO/IEC 17025-2017 and ANSI Z-540-1 using laboratory standards which are traceable to the SI International System of Quantities through the National Institute of Standards and Technology (NIST), and or other Accredited bodies except where none exist. Tests are carried out in environmental conditions controlled to the extent appropriate to the instrument's specification. This certificate shall not be reproduced except in full without the written approval of the laboratory. The uncertainty results meet the requirements of the ISO/IEC 17025-2017 standard and ILAC Doc.P14. Statements of conformity (e.g. Pass or Fail) are made in accordance with Simple Acceptance decision rules as defined in ILAC G8 with a TUR of 4:1 or greater. The customer is responsible for considering whether the inclusion of the uncertainties shown on the certificate would prevent their use of the equipment based on their risk evaluations. Results are accredited unless annotated with an asterisk "\*". The results presented are only applicable to the Model/Serial number shown.

Ambient Conditions of Laboratory

Temperature (°C): **26.3**Relative Humidity (%): **38.0**

Technician:

**Trevor Einspahr**

Technician Signature: \_\_\_\_\_



Calibration Equipment				
Model	Description	Serial Number	Certificate	Due Date
ECO-2G	Oscilloscope	SDS62CCD5R0146	5523631031001725	6/17/2025
CST2	ESD Target and Attenuator	003170	A25011501RY	1/15/2026
ESVM	Electrostatic Voltmeter	20200105-01	A24062101TE	6/21/2025

Calibration method used: IEC 61000-4-2:Ed2

Contact discharge: Peak Voltage / Zsys (0.068848) = Peak Current

Condition as found:	In tolerance
Condition as left:	In tolerance

Measuring Uncertainties	
ESD Voltage	± 2.1%
ESD Current	± 2.8%
Rise Time	± 3.2%
Current at 30ns and 60ns	± 2.8%

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

IEC 61000-4-2 Ed.2 defines five discharges per measured parameter. Each of the pulses must individually fulfill the ESD pulse specification. The shown measurements of peak and rise time are the mean value calculated from all pulses.



### Air Discharge DC Output Voltage

*IEC Specification:  $\pm 5\%$  of Nominal Value*

IEC-Level	Nominal Voltage (kV)	Measured Voltage (kV)	Deviation	Verdict
Pos. 1	2.00	<b>1.96</b>	2.0%	PASS
Pos. 2	4.00	<b>3.99</b>	0.2%	PASS
Pos. X	6.00	<b>5.99</b>	0.2%	PASS
Pos. 3	8.00	<b>7.97</b>	0.4%	PASS
Pos. 4	15.00	<b>15.13</b>	0.9%	PASS
Pos. X	25.00	<b>24.63</b>	1.5%	PASS
Pos. X	30.00	<b>29.43</b>	1.9%	PASS
Neg. 1	-2.00	<b>-2.02</b>	1.0%	PASS
Neg. 2	-4.00	<b>-4.03</b>	0.8%	PASS
Neg. X	-6.00	<b>-6.03</b>	0.5%	PASS
Neg. 3	-8.00	<b>-8.02</b>	0.2%	PASS
Neg. 4	-15.00	<b>-14.84</b>	1.1%	PASS
Neg. X	-25.00	<b>-24.84</b>	0.6%	PASS
Neg. X	-30.00	<b>-29.76</b>	0.8%	PASS



## Contact Discharge Pulse Measurements

### Peak Current

*IEC Specification:  $\pm 15\%$  of Nominal Value*

*Average Peak Voltage / Zsys (0.068848) = Average Peak Current*

IEC- Level	Test Voltage (kV)	Nominal Current (A)	Measured Voltage (V)	Peak Current (A)	Deviation	Verdict
Pos. 1	2.00	7.50	<b>0.497</b>	7.22	3.7%	PASS
Pos. 2	4.00	15.00	<b>1.03</b>	15.00	0.0%	PASS
Pos. 3	6.00	22.50	<b>1.54</b>	22.40	0.5%	PASS
Pos. 4	8.00	30.00	<b>2.05</b>	29.82	0.6%	PASS
Pos. X	15.00	56.25	<b>3.92</b>	56.95	1.2%	PASS
Pos. X	25.00	93.75	<b>6.40</b>	92.94	0.9%	PASS
Neg. 1	-2.00	-7.50	<b>-0.500</b>	-7.26	3.2%	PASS
Neg. 2	-4.00	-15.00	<b>-1.03</b>	-14.96	0.3%	PASS
Neg. 3	-6.00	-22.50	<b>-1.55</b>	-22.56	0.3%	PASS
Neg. 4	-8.00	-30.00	<b>-2.09</b>	-30.28	0.9%	PASS
Neg. X	-15.00	-56.25	<b>-3.83</b>	-55.62	1.1%	PASS
Neg. X	-25.00	-93.75	<b>-6.32</b>	-91.77	2.1%	PASS

### Risetime

*IEC Specification: 800ps  $\pm 25\%$*

IEC- Level	Test Voltage (kV)	Measured Risetime (ps)	Deviation	Verdict
Pos. 1	2.00	<b>822</b>	2.8%	PASS
Pos. 2	4.00	<b>818</b>	2.3%	PASS
Pos. 3	6.00	<b>840</b>	5.0%	PASS
Pos. 4	8.00	<b>825</b>	3.1%	PASS
Pos. X	15.00	<b>847</b>	5.9%	PASS
Pos. X	25.00	<b>845</b>	5.6%	PASS
Neg. 1	-2.00	<b>815</b>	1.9%	PASS
Neg. 2	-4.00	<b>835</b>	4.4%	PASS
Neg. 3	-6.00	<b>843</b>	5.4%	PASS
Neg. 4	-8.00	<b>853</b>	6.6%	PASS
Neg. X	-15.00	<b>843</b>	5.4%	PASS
Neg. X	-25.00	<b>859</b>	7.4%	PASS



### Curve Decay Points at 30 and 60 ns

*IEC Specification:  $\pm 30\%$  of Nominal Value*

*Average Peak Voltage / Zsys (0.068848) = Average Peak Current*

**30ns**

IEC- Level	Test Voltage (kV)	Nominal Current (A)	Measured Voltage (V)	Peak Current (A)	Deviation	Verdict
Pos. 1	2.00	4.00	<b>0.265</b>	3.85	3.8%	PASS
Pos. 2	4.00	8.00	<b>0.566</b>	8.22	2.8%	PASS
Pos. 3	6.00	12.00	<b>0.850</b>	12.35	2.9%	PASS
Pos. 4	8.00	16.00	<b>1.13</b>	16.35	2.2%	PASS
Pos. X	15.00	30.00	<b>2.23</b>	32.38	7.9%	PASS
Pos. X	25.00	50.00	<b>3.79</b>	55.05	10.1%	PASS
Neg. 1	-2.00	-4.00	<b>-0.279</b>	-4.05	1.3%	PASS
Neg. 2	-4.00	-8.00	<b>-0.581</b>	-8.44	5.5%	PASS
Neg. 3	-6.00	-12.00	<b>-0.869</b>	-12.62	5.2%	PASS
Neg. 4	-8.00	-16.00	<b>-1.17</b>	-17.01	6.3%	PASS
Neg. X	-15.00	-30.00	<b>-2.19</b>	-31.81	6.0%	PASS
Neg. X	-25.00	-50.00	<b>-3.73</b>	-54.18	8.4%	PASS

*IEC Specification:  $\pm 30\%$  of Nominal Value*


*Average Peak Voltage / Zsys (0.068848) = Average Peak Current*

**60ns**


IEC- Level	Test Voltage (kV)	Nominal Current (A)	Measured Voltage (V)	Peak Current (A)	Deviation	Verdict
Pos. 1	2.00	2.00	<b>0.137</b>	1.99	0.7%	PASS
Pos. 2	4.00	4.00	<b>0.285</b>	4.14	3.5%	PASS
Pos. 3	6.00	6.00	<b>0.446</b>	6.48	8.0%	PASS
Pos. 4	8.00	8.00	<b>0.566</b>	8.22	2.8%	PASS
Pos. X	15.00	15.00	<b>1.14</b>	16.62	10.8%	PASS
Pos. X	25.00	25.00	<b>1.94</b>	28.18	12.7%	PASS
Neg. 1	-2.00	-2.00	<b>-0.135</b>	-1.96	2.0%	PASS
Neg. 2	-4.00	-4.00	<b>-0.278</b>	-4.04	0.9%	PASS
Neg. 3	-6.00	-6.00	<b>-0.430</b>	-6.25	4.1%	PASS
Neg. 4	-8.00	-8.00	<b>-0.566</b>	-8.22	2.8%	PASS
Neg. X	-15.00	-15.00	<b>-1.06</b>	-15.38	2.5%	PASS
Neg. X	-25.00	-25.00	<b>-1.82</b>	-26.44	5.7%	PASS



2kV




-2kV






4kV




-4kV








6kV




-6kV








8kV




-8kV






15kV




-15kV






25kV



-25kV



End of Calibration Report: ESD Simulator 30kV 150pF/330Ω cert. rev. 2